



ИЦ ГЛОБАЛТЕСТ ООД

1618 София, ул. "Крушовски връх" 31
тел./факс (02) 955 9862; 955 9837; 44 10 329
www.globaltest-bg.com - office@globaltest-bg.com

TC GLOBALTEST Ltd.

31, Krushovski vrah Str, 1618 Sofia, Bulgaria
tel./fax (+359 2) 955 9862; 955 9837; 44 10 329
www.globaltest-bg.com - office@globaltest-bg.com

REPORT OF ANALYSES

Number / Date: RN17484 / 31.10.2018

Application for testing No.: RN17484/ 31.10.2018

Period of testing: 23.10.2018-30.10.2018

Applicant:	Dragonfly Biosciences Bulgaria Ltd
Applicant address:	9 Vladimir Vazov Blvd., Office 8, entrance 1, fl.4, 1510 Sofia
Appl. letter No. and date:	-
Arrival of samples:	Personally
Date of arrival of the sample:	23.10.2018
Tested object:	11,1% CBD P.C. Oil 1000mg; Batch No: 0810
Producer:	Dragonfly Biosciences Bulgaria
Standards and regulations:	Ph. Eur 2.6.13, Ph. Eur. 2.6.13, Ph.Eur. 2.6.12, Ph.Eur. 2.6.13
Number of samples:	1 count (sterile syringe)
Analyses performed by:	Biotechnological engineer Sonya Balkanska



TC GLOBALTEST Manager: Dimitar Tanev, chemical engineer

Report of analyses composed by: Boyko Pekanov

stamp

signature

The results of analyses are listed on pages 2 through 2

NOTE: The results of analyses refer to the tested sample only. Excerpts and/or copies of the Report may not be made without the written consent of the test laboratory.

Reports of analyses are valid only with a "wet" seal or signed electronically.

The tested samples are not subject to return to the applicant due to the destructive nature of the tests performed.

RESULTS OF ANALYSIS - Report of Analysis RN17484/ 31.10.2018

Characteristic	Method	Meas. un.	Norm	Result	Uncertainty	Measuring Conditions
Bile gram negative bacteria	Ph. Eur 2.6.13	-	Not specified	Absence	-	35.0± 0.1 °C
Escherichia coli	Ph. Eur. 2.6.13	-	Not specified	Absence	-	35.0± 0.1 °C
Salmonella	Ph. Eur. 2.6.13	-	Not specified	Absence	-	35.0± 0.1 °C
Staphylococcus aureus	Ph.Eur. 2.6.13	-	Not specified	Absence	-	3050± 0.1 °C
Total aerobic microbial count (TAMC)	Ph.Eur. 2.6.12	cfu/ml	Not specified	Absence	-	35.0± 0.1 °C
Total combined yeast and mould count (TYMC)	Ph.Eur. 2.6.12	cfu/ml	Not specified	Absence	-	25± 0.1 °C

Quality assurance Yana Georgieva, chemical engineer